

## Production Description

Name	Isolation Coveralls with Hood
Material	PP+PE Coating
Weight	50-65GSM
MOQ	10,000PCS
Packing	Normally,1pcs/polybag, 50pcs/ctn
Delivery	Within 2 weeks
features	Breathable coverall fabric Front long zipper,elastic cuffs/waist and attached hood

### \*Application

COMMON USES in clinics, hospitals, laboratories, workshops,etc. For general isolation and protection.

### \*Safety and Anti-Epidemic Medical-Grade Fabric

Lightweight multi-ply material that is fluid-resistant and breathable

Safe and strong enough for personal protection

Gowns are generously sized for full coverage

With elastic or knit cuffs

\*



碧罗

BECOLOR

Protective Coveralls with Hood

*\*Effectively prevent  
the invasion of bacteria*



New  
Material

Safe  
Soft

Dust  
proof



**ASIANGREAT**  上海亚厚纺织科技有限公司  
Shanghai Asiangreat Textile Technology Co., Ltd







180000122582



中国认可  
国际互认  
检测  
TESTING  
CNAS L2278

# 检测报告

武汉海关技术中心





# 武汉海关技术中心

正本

Technology Center Of Wuhan Customs District

## 检测报告

Test Report

申报样品名称: 一次性隔离衣(非医用)  
Name of sample declared: Disposable isolation gown(non-medical use)

报告编号(No.of Series): 21202000746-1  
报告日期(Date): 2020-05-15  
报告页数(Page): 共1页 第1页

委托人: 湖北君橙服饰有限公司  
Consignor: HUBEI ORANGE FASHION CO.,LTD

接样日期: 2020-05-12

Date of reception:

样品来源: 委托人送样  
Origin of sample: Delivering by Consignor

检测依据: GB 15979-2002 附录 B

Method of inspection:

检测项目及结果:

Item inspected and its data:

样品标记 Sample Mark	2#
样品编号 Sample No.	202003069
细菌菌落总数 Total bacterial count (CFU/g)	84
大肠菌群 Coliforms	未检出 ND

\*\*\*\*\*

以下空白  
BLANK BELOW



管孝子

签署(Signature):

印章(Seal):

Notice:1.本报告仅对到样负责。(It is only responsible for the sample received.)

2.未加盖检验检测报告专用章无效。(It is invalid without original seal of inspection & test report.)

3.未经本中心书面许可,不得部分复制报告。(It can not be copied separately without authorized permission.)

地址: 湖北省武汉市汉阳琴台大道 588 号

电话(Tel): 027-58905910; 58905912

Add.: No.588Qin Tai Avenue, Hanyang, Wuhan, Hubei Province of P.R.C

传真(Fax): 027-58905911

**产品合格证**  
Certificate of Conformity

产品名称	聚丙烯酰胺 (PAM)
规格	2000-05-12
生产厂家	大庆市大同区大同镇大同村大同队
生产日期	2020-05-12
产品数量	1
检验员	张明
检验日期	2020-05-12
检验地点	大庆市大同区大同镇大同村大同队
检验结果	合格

4 2 0 2 0 0 3 0 6 9 4

批号: 2120200746

品名: 聚丙烯酰胺 (PAM)

规格: 大庆市大同区大同镇大同村大同队

生产日期: 2020-05-12

产品数量: 1

检验员: 张明

检验日期: 2020-05-12

检验地点: 大庆市大同区大同镇大同村大同队

STORERS DISTRICT  
中心站



## 检验检测报告

(电子版)

 防伪查询网址: [www.gttc.net.cn](http://www.gttc.net.cn)

防伪码: OEIM-8603-54

共3页 第1页


扫码下载报告



No: 200128094

委托单位	武汉宝齐莱商贸有限公司 地址: 湖北省武汉市江汉区淮海路泛海国际SOHO城8号楼1506室		
客户认定信息	一次性隔离衣(非医用) 4件 商标: BECOLOR 批号款号: 批号: 20200516 款号: CF23156 颜色: 浅蓝 材质: 100%聚丙烯 号型规格: OFSPC-XL 生产单位: 湖北君橙服饰有限公司		
检验性质	委托检测	样品受理/测试开始日期	2020-05-17
		报告签发日期	2020-05-23
判定依据	Q/OFGLY01-2020 《一次性隔离衣(非医用)》		
综合检验结论	---		
检验检测结果	检验检测项目	判定依据	判定
	断裂强力	Q/OFGLY01-2020	符合
	断裂伸长率	Q/OFGLY01-2020	符合
	抗渗水性	Q/OFGLY01-2020	符合
备注	本报告中检验检测项目均在相应标准规定的环境条件下进行(有注明的除外)。 复印件、副本未重新加盖报告书确认章无效。 本报告检验检测地址为广州市番禺区珠江路1号。		

签发: 刘圆 工程师






样品图片 (电子版)

No:200128094

共3页 第2页



## 检验检测报告附页 (电子版)

No: 200128094

共3页 第3页

检验检测项目 (计量单位) [样品识别]	测试方法	标准值及允差	检验检测结果	判定	备注
● 断裂强力(N)	GB/T 3923.1-2013 夹距:100mm 速度:100mm/min	纵向 $\geq 10$ 横向 $\geq 10$	纵向 49 横向 95	符合	注
● 断裂伸长率(%)	GB/T 3923.1-2013 夹距:100mm 速度:100mm/min	纵向 $\geq 15$ 横向 $\geq 15$	纵向 101 横向 115	符合	注
● 抗渗水性 (cmH <sub>2</sub> O)	GB/T 4744-2013 增压法:60cmH <sub>2</sub> O/min 水温:20.0℃ 正面向下接触水面	$\geq 10$	1# 71.3 2# 73.9 3# 36.4 4# 64.5 5# 68.6 平均值 62.9	符合	注
备 注	注: 判定指标来源于客户要求的企业标准(Q/OFGLY01-2020)。我单位获CMA认定授权的检验能力覆盖Q/OFGLY01-2020所检项目。				



——本报告结束——

